

VERIFICATION OF CALIBRATION REPORT
of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130129</u> ✓	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>JAMES MICHELS</u>	ID# <u>6502</u> Date <u>3/11/10</u> ✓
A Agency <u>NORTH SLOPE BOROUGH POLICE</u> Phone # <u>852-6111</u>	
Instrument Location <u>BARROW - SQUAD ROOM</u>	
B Alco S/N <u>X301568</u> Target Value <u>.084</u> High Pressure <u>900</u>	
Alco Test Values <u>.083</u> ✓ <u>.083</u> ✓ 1 st Alco 2 nd Alco	
Signature <u>[Signature]</u> ✓	
(OVER)	

BMB
3/18/2010

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 2nd day of April, 2010.

[Signature] (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130129 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130129
MARCH 11, 2010 ✓ *BMB 3/11/2010*
OPERATOR'S NAME:
MICHAEL JONES/A
OPERATOR'S PHONE: 606
SUBJECT'S LAST NAME:
VOC
SUBJECT'S FIRST NAME/MI :
VOC
A.L. # VOC
DEPT/AGENCY: DMV
CASE/REPORT: VOC
TEST TYPE: V
ALCO TARGET VALUE: .004 ✓
ALCO S/N: X301568 ✓

.004 ADJUSTED FOR 29.59 in
ALCO TARGET .003 18:00
BLANK TEST .000 18:00
INTERNAL STANDARD VERIFIED 18:00
ALCO TV 29.59 in .003 18:00
BLANK TEST .000 18:00
SUBJECT SAMPLE *BMB 3/11/2010* .000 18:00
BLANK TEST .000 18:00
ALCO TV 29.59 in .003 18:00
BLANK TEST .000 18:00

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130129 ✓
MARCH 11, 2010 ✓
TIME 18:06
— DIAGNOSTIC CHECK —
COMPUTER: OKY
PROGRAM: OKY
SOFTWARE DATE: 01/09/09
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 29.59 in
FLOW DETECTOR: OKY
PUMP
HIGH SPEED: OKY
DETECTOR: OKY
FILTERS: OKY
GURTZ STANDARD: OKY
CALIBRATION: OKY
PRINTER TEST
! " # \$ % & ' () * + , - . / : ; < = > ? @ [\] ^ _ ` { | } ~ ¡ ¢ £ ¤ ¥ ¦ § ¨ © ª « ¬ ® ¯ ° ± ² ³ ´ µ ¶ · ¸ ¹ º » ¼ ½ ¾ ¿ À Á Â Ã Ä Å Æ Ç È É Ê Ë Ì Í Î Ï Ñ Ò Ó Ô Õ Ö × Ø Ù Ú Û Ü Ý Þ ß à á â ã